

T2

Dead volumes requirements based on sample cup size for common instrument platforms

| Instrument | Sample cup size (mL) | Dead volume (μL) | Assay Volume (μL) |
|------------------------|----------------------|-------------------------------|--------------------------------|
| Beckman Coulter® Au480 | 0.5 | 40 | 1-25 |
| | 2.0 | 360 | |
| Roche Cobas® 6000 | 1.5 | 50 | 1-35 |
| | 2.5 | 100 | |